

CERTIFICATE

TYPE TEST OF ENERGY METERS

Applicant : Bluestar Electrical Meter Research Institute
Contact person : Changchun Yang
Address : 181 Hanzhongmen Street
Postal code, Place : 210036210036 Nanjing
Country : China

Manufacturer : Microstar Electric Company Limited
Address : 10 Dajiang Road
Postal code, Place : 210039 Nanjing
Country : China

Instrument : Electronic single-phase two-wire energy meter
direct connected

Type : C2000
Register : LCD
Accuracy Class : 1/ B
Measurement range : 240 V
10(100) A
50 Hz
500 imp./kWh

Use : Indoor
Protection Class : II
Registry method : programmable

The energy meter meets the requirements of:


IEC 62052-11 (2003) : Electricity metering equipment (a.c.) - General requirements, tests and test conditions -
Metering equipment
IEC 62053-21 (2003) : Electricity metering equipment (a.c.) - Static meters for active energy (classes 1 and 2)

Based on a non-recurrent examination.

The results are recorded in our test report with reference 70068157-TDT 6378-10.

KEMA Quality B.V.
Arnhem, October 1, 2010


A.P.M. Baars
Certification manager


ir. P.J.J.G. Nabuurs
Managing Director
Certificate nr. 70068157-TDT 6378-10

The investigation reported here does not confer any right to use an approbation mark granted by KEMA.

Integral publication of this certificate and adjoining reports is allowed.
This certificate is issued provided that neither KEMA nor the RvA assumes any liability.